

DIETHYSTILBESTROL ELISA TEST KIT
MANUAL

Diethylstilbestrol ELISA Test Kit

Catalogue Number. IP100042

Principle

This test kit is based on the competitive enzyme immunoassay for the detection of Diethylstilbestrol in the feed, urine, liver, meat, shrimp and fish. The conjugate antigen is precoated on the microwell stripes. The Diethylstilbestrol in the sample competes with the conjugate antigen pre-coated on the microwell stripes, to interact with the antibodies against Diethylstilbestrol. After the addition of the TMB substrate is added for coloration. The optical density (OD) value of the sample has a negative correlation with the content of Diethylstilbestrol in the sample. This value is compared to the standard curve and the content of the corresponding Diethylstilbestrol is subsequently obtained.

Technical specifications

Sensitivity: 0.1 ppb

Incubation Temperature: 37°C

Incubation Time: 30min—30min—15min

Detection limit: Tissue (Shrimp, fish) 0.2 ppb, Pork/liver, chicken/liver 2 ppb, Muscle tissue(Method Two) 0.1ppb, Urine 0.6 ppb, Feed 20 ppb

Cross-reactions rate: DES 100%, Dienestrol 38.5%, Hexestrol 8.5%, Ethinylestradiol < 0.1%, Estriol < 0.1%

Recovery rate: Urine 70 ±10%, Feed 90 ±10%, Tissue 85 ±10%

Components

1	Micro-well strips	12 strips with 8 removable wells each	
2	6× standard solution (1mL each)	0ppb, 0.1ppb, 0.3ppb, 0.9ppb, 2.7ppb and 8.1ppb	
3	Enzyme conjugate	12ml	red cap
4	Antibody working solution	7ml	blue cap
5	Substrate A	7ml	white cap
6	Substrate B	7ml	black cap
7	Stop solution	7ml	yellow cap
8	20× concentrated washing buffer	40ml	white cap
9	5× concentrated redissolving solution	50ml	transparent cap

Materials required but not provided

Equipments: microplate reader, printer, homogenizer, nitrogen-drying device, vortex, centrifuge, measuring pipets, balance (a sensibility reciprocal of 0.01 g)

Micropipettors: single-channel 20-200 μL, 100-1000 μL, and multi-channel 30~300 μL;

Reagents: NaOH, Acetonitrile (CH₃CN), Acetone, deionized water, H₃PO₄ (85%), ethyl acetate, CHCl₃

Sample pre-treatment

Instructions

- 1) Only the disposable tips can be used for the experiments and the tips must be changed when used for absorbing different reagents;
- 2) Before the experiment, each experimental utensil must be clean and should be re-cleaned if necessary, in order to avoid the contamination that interferes with the experimental results.

Solution preparation before sample pre-treatment

- 1) 6 M H₃PO₄ : dissolve 100 mL H₃PO₄ (85%) in 150 mL deionized water, mix properly
- 2) 1 M NaOH : dissolve 4 g NaOH in deionized water to 100 mL
- 3) 2 M NaOH : dissolve 8 g NaOH in deionized water to 100 mL
- 4) Acetonitrile- Acetone: add 80 mL Acetonitrile and 20 mL Acetone, mix evenly
- 5) The 5×concentrated redissolving solution is mixed with deionized water at 1:4 (1 mL concentrated redissolving solution + 4 mL deionized water), used for the treated sample redissolving

Samples preparation

a) Tissue (Chicken, duck, pork/liver, shrimp, fish)

1. Weigh 2 ± 0.05 g of the homogenized sample, add 6 mL Acetonitrile-Acetone, shake for 2 min, and centrifuge at above 4000 r/min at 15 °C for 10 min.
2. Transfer 3 mL supernatant into a new centrifuge tube, blow to dry with nitrogen or air at 60 °C.
3. Add 0.5 mL CHCl₃, vortex for 20 sec, add 2 mL 2 M NaOH, vortex for 30 sec, centrifuge at above 4000 r/min for 5 min.
4. Take 1 mL supernatant, add 200 μL 6 M H₃PO₄, vortex for 5 sec.
5. Add 3 mL Acetonitrile (CH₃CN) for extraction, shake properly for 2 min, centrifuge at above 4000 r/min at room temperature (20-25 °C) for 10 min, take the upper layer, blow to dry with nitrogen or air at 60 °C.
6. Dissolve dry residues in 1 mL of the diluted redissolving solution.

Dilute as following method for different samples

- 1) Shrimp and fish-----directly take 50 μL water phase for detection;
- 2) Pork/liver, Chicken/liver-----

take 50 μL water phase, add 450 μL of the diluted redissolving solution, shake properly for 30s. Take 50 μL for analysis.

Fold of dilution of the sample:

Shrimp and fish----2 Pork/liver, Chicken/liver-----20

b) Muscle tissue

Method Two

- 1) Take 2 ± 0.05 g homogenized sample, add 2ml 2M NaOH, shake for 2min;
- 2) Add 8ml ethyl acetate, shake violently for 5min;
- 3) Centrifuge at above 4000 r/min at 15 °C for 10 min;
- 4) Take 4ml upper organic phase, blow to dry with nitrogen or air at 60 °C
- 5) Dissolve dry residues in 1 mL of the diluted redissolving solution, shake for 30s.

Fold of dilution of the sample: 1

c) Feed

- 1) Weigh 2 ± 0.05 g of the homogenized sample, add 8 mL Acetonitrile, shake properly for 2 min, centrifuge at above 4000 r/min at 15 °C for 10 min

- 2) Take 2 mL supernatant into a new centrifuge tube, blow to dry with nitrogen or air at 60 °C.
- 3) Add 0.5 mL CHCl₃, vortex for 20 sec, add 2 mL 1 M NaOH, vortex for 30 sec, centrifuge at above 4000 r/min for 5 min.
- 4) Take 1 mL supernatant, add 100 µL 6 M H₃PO₄, vortex for 5 sec
Dilute as following method for different samples
Compound feed--
take 50 µL sample, add 950 µL of the diluted redissolving solution, vortex for 30 sec, take 50 µL for analysis;
Concentrated /Premixed feed--
take 25 µL sample, add 975 µL of the diluted redissolving solution, vortex for 30 sec, take 50 µL for analysis.
Fold of dilution of the sample:
Compound feed -----100 Concentrated / Premixed feed -----200

d) Urine

1. Take 2 mL urine into centrifuge tube, centrifuge at above 4000 r/min at room temperature (20-25) for 10 min, stop when it is clear.
2. Transfer 1 mL clear urine into centrifuge tube, add 1 mL 1 M NaOH, shake vigorously for 5 min
3. Add 100 µL 6 M H₃PO₄, vortex for 30 sec
4. Add 8 mL CHCl₃ for extraction, shake properly for 5 min, centrifuge at above 4000 r/min at 15°C for 10 min.
5. Remove the upper layer (water phase), take 4 mL of the lower layer, blow to dry with nitrogen or air at 60 °C.
6. Dissolve dry residues in 3 mL of the diluted redissolving solution, shake for 30s.
7. Take 50 µL for analysis.
Fold of dilution of the sample: 6

ELISA procedures

Instructions

1. Bring all reagents and micro-well strips to the room temperature (20-25 °C) before use.
2. Return all reagents to 2-8 °C immediately after use
3. The reproducibility of the ELISA analysis, to a large degree, depends on the consistency of plate washing. The correct operation of plate washing is the key point in the procedures of ELISA;
4. For the incubation at constant temperatures, all the samples and reagents must avoid light exposure, and each microplate should be sealed by the cover membrane.

Operation procedures

1. Take out the kit from the refrigerated environment. Take out all the necessary reagents from the kit and place at the room temperature (20-25°C) for at least 30 min. Note that each liquid reagent must be shaken to mix evenly before use
2. Take the required micro-well strips and plate frames. Re-sealed the unused microplate, stored at 2-8 °C, not frozen
3. Solution preparation: dilute the 20× concentrated washing buffer with the deionized water to 800 mL (or just to the required volume) for use
4. Numbering: number the micro-wells according to samples and standard solution; each sample and standard solution should be performed in duplicate; record their positions.
5. Add 50 µL of the sample or standard solution to separate duplicate wells; add 50 µL of the antibody

working solution into each well. Seal the microplate with the cover membrane, and incubate at 37°C for 30 min.

6. Pour the liquid out of the microwells, add 250 µL/well of washing buffer for 10 sec, repeat four to five times, then flap to dry with absorbent paper (if there are the bubbles after flapping, cut them with the clean tips).

7. Add 100 µL enzyme conjugate into every well, seal the microplate with the cover membrane, react at 37°C for 30 min, continue as described in 6

8. Coloration: add 50 µL of the substrate A solution and then 50 µL of the B solution into each well. Mix gently by shaking the plate manually, seal the microplate with the cover membrane and incubate at 37 °C for 15 min at dark for coloration.

9. Determination: add 50 µL of the stop solution into each well. Mix gently by shaking the plate manually. Set the wavelength of the microplate reader at 450 nm to determine the OD value (Recommend to read the OD value at the dual-wavelength 450/630 nm within 5 min)

Result judgment

There are two methods to judge the results; the first one is the rough judgment, while the second is the quantitative determination. Note that the OD value of the sample has a negative correlation with the content of Diethylstilbestrol (DES).

Qualitative determination

The concentration range (ng/mL) can be obtained from the comparison the average OD value of the sample with that of the standard solution. Assuming that the OD value of the sample I is 0.3, and that of the sample II is 1.0, while those of the standard solutions are as the followings: 2.243 for 0 ppb, 1.816 for 0.1 ppb, 1.415 for 0.3 ppb, 0.74 for 0.9 ppb, 0.313 for 2.7 ppb and 0.155 for 8.1 ppb, accordingly the concentration range of the sample is 2.7 to 8.1ppb, and that of the sample II is 0.3 to 0.9ppb. (multiplied by corresponding dilution fold)

Quantitative determination

The mean values of the absorbance values obtained for the percentage of the average OD value (B) of the sample and the standard solution divided by the OD value (B₀) of the first standard solution (0 standard) and subsequently multiplied by 100%, that is,

$$\text{Percentage of absorbance value} = (B / B_0) \times 100\%$$

B—the average OD value of the sample or the standard solution

B₀—the average OD value of the 0 ng/mL standard solution

Draw the standard curve with the absorption percentages of the standard solutions and the semi-logarithm values of the Diethylstilbestrol(DES) standard solutions (ng/mL) as Y- and X-axis, respectively. Read the corresponding concentration of the sample from the standard curve by incorporating its absorption percentage into the standard curve. The resulting value is subsequently multiplied by the corresponding dilution fold, thus finally obtaining the Diethylstilbestrol (DES) concentration in the sample.

Using the professional analyzing software of this kit will be more convenient for the accurate and rapid analysis of a large amount of samples. (Please contact us for this software)

Precautions

1. The room temperature below 25 °C or the temperature of the reagents and the samples being not returned to the room temperature (20-25 °C) will lead to a lower standard OD value

2. Dryness of the microplate in the washing process will be accompanied by the situations including the non-linear standard curves and the undesirable reproducibility; So continue to next step immediately after washing.
3. Mix evenly, otherwise there will be the undesirable reproducibility
4. The stop solution is the 2 M sulfuric acid solution, avoid contacting with the skin;
5. Do not use the kit exceeding its expiry date. The use of diluted or adulterated reagents from the kits will lead to the changes in the sensitivity and the detecting OD values. Do not exchange the reagents from the kits of different lot numbers to use
6. Put the unused microplate into an auto-sealing bag to re seal it. The standard substance and the colorless color former is light sensitive, and thus they cannot be directly exposed to the light
7. Discard the coloration solution with any color that indicates the degeneration of this solution. The detecting value of the 0 standard solution of less than 0.5 (A450 nm < 0.5) indicates its degeneration
8. The optimum reaction temperature is 37 °C, and too high or low temperatures will result in the change in the detecting sensitivity and OD values.

Storage: store at 2-8 °C, not frozen.

Expiry date: 12 months; date of production is on box.